



EC-Conformity Declaration

For the following equipment :

Product Name: Switching Power Supply

Model Designation:

GEM60lbzzzzz,GSM60Ebzzzzz (b=05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A; zzzzz=0-9, A-Z hyphen or blank)

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were applied :

RoHS Directive (2011/65/EU)

MDD Directive (93/42/EEC) :

EN 60601-1:2006+A11+A1+A12

TUV Certificate No : TA 50422408

EMI (Electro-Magnetic Interference)

Conducted / Radiated emission	EN55011:2009+A1:2010(Group 1)	CLASS B
Harmonic current	EN61000-3-2:2014	CLASS A
Voltage flicker	EN61000-3-3:2013	

EMS (Electro-Magnetic Susceptibility)

EN60601-1-2:2015

ESD air	EN61000-4-2:2009	Level 4	15KV
ESD contact	EN61000-4-2:2009	Level 4	8KV
RF field susceptibility	EN61000-4-3:2010	Level 3	10V/m (80MHz~2.7GH) Level x 9~28V/m(385MHz~5.78GHz)
EFT bursts	EN61000-4-4:2012	Level 3	2KV
Surge susceptibility	EN61000-4-5:2014	Level 3	1KV / L-N
Conducted susceptibility	EN61000-4-6:2014	Level 2	3 Vrms (0.15~80MHz) Level x 6 Vrms (6.76MHz~54MHz)
Magnetic field immunity	EN61000-4-8:2010	Level 4	30A/m
Voltage dip, interruption	EN61000-4-11:2004+A1:2017	100% dip 1 periods 30% dip 25 periods 100% interruptions 250 periods	

Note:

The power supply is considered as a component that will be operated in combination with final equipment. Since EMC performance will be affected by the complete system, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again For guidance on how to perform these EMC tests, please refer to TDF (Technical Documentation File).

This Declaration is effective from serial number EJ xxxxxx B1848R

Person responsible for marking this declaration :

Mean Well Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan

(Manufacturer Address)

Alex Tsai /Director, , Marketing Department :

(Name / Position)

(Signature)

Taiwan

(Place)

Nov.23rd.2018

(Date)