



Declaration of Conformity

| For the following ed | quipment | : |
|----------------------|----------|---|
|----------------------|----------|---|

Product Name: Medical Type Switching Power Supply

Model Designation: MSP-600-x (x=3.3,5,7.5,12,15,24,36,48)

is herewith confirmed to comply with the requirements set out in the Council Directive 93/42/EEC concerning Medical devices, the following standards were applied:

RoHS Directive (2011/65/EU), (EU)2015/863

MDR Directive (EU) 2017/745

EN 60601-1-2:2015+A1:2021

EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

| | EN 55011:2016/A2:2021 (Group 1) | | Class B | | | | |
|---------------------------------------|---------------------------------|---------|--------------------------|--|--|--|--|
| Harmonic current | EN IEC 61000-3-2:2019 | | | | | | |
| Voltage flicker | EN 61000-3-3:2013/A1:2019 | | | | | | |
| EMS (Electro-Magnetic Susceptibility) | | | | | | | |
| ESD air | EN 61000-4-2:2009 | Level 4 | 15KV | | | | |
| ESD contact | EN 61000-4-2:2009 | Level 4 | 8KV | | | | |
| RF field susceptibility | EN IEC 61000-4-3:2020 | Level 3 | 10V/m(80MHz-2.7GHz) | | | | |
| RF field susceptibility | EN IEC 61000-4-3:2020 | Table 9 | 9~28V/m (385MHz~5.78GHz) | | | | |
| EFT bursts | EN 61000-4-4:2012 | Level 3 | 2KV/100KHz | | | | |

| RF field susceptibility | EN IEC 61000-4-3:2020 | Table 9 | 9~28V/m (385MHz~5.78GHz) | |
|--------------------------|--------------------------|------------------------------|--------------------------|-----------------------|
| EFT bursts | EN 61000-4-4:2012 | Level 3 | 2KV/100KHz | |
| Surge susceptibility | EN 61000-4-5:2014/A1:201 | 7 Level 4 | 2KV/Line-Line | 4KV/Line-Earth |
| Conducted susceptibility | EN 61000-4-6:2014 | Level 3 | 10V | |
| Magnetic field immunity | EN 61000-4-8:2010 | Level 4 | 30A/m | |
| Voltage din interruption | EN IEC 61000-4-11:2020 | 100% din 1 neriods 30% din 2 | 25 neriods 100% inte | rruntions 250 periods |

Note:

A component power supply with load will be installed into final equipment which consists of an electronically shielded metal enclosure. Since EMC performance will be affected by the complete installation, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again.

The EMC tests mentioned above are performed using a well defined metal plate to simulate said metal enclosure. For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies".(as available on http://www.meanwell.com)".

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC4xxxxxxx

Person responsible for marking this declaration:

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Djst., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D:

(Name / Position)

(Signature)

Alex Tsai/Director, Product Strategy Center:

(Name / Position)

(Signature)

Taiwan (Place) Jan. 30th, 2024 (Date)