

# Certificate



Certificate no.

TA 50393494 01

**License Holder:**

MEAN WELL Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248  
Taiwan

**Manufacturing Plant:**

MEAN WELL Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248  
Taiwan

Test report no.: ZTW1- 50095166 001

Client Reference: FL-01082017-NMS240

Tested to: IEC 60601-1:2005+A1  
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

Type Designation : NMS-240-05 (MW) 10  
Output Voltage : DC 5V  
Output Current : 36A  
max. Ambient Temperature : 50°C  
max. Operational Altitude: 5000m  
Protection Class : I

- Remarks: 1) The risk management and usability requirements of the standard were not addressed;  
2) The power modules under evaluation shall be equipped exclusively in power supply unit manufactured by MEANWELL type either NMP-1K2 series or NMP-650 series;  
3) This certificate is valid for maximum 5 years.

Appendix: 1

10

Licensed Test mark:



TÜV Rheinland Taiwan Ltd  
Signature

Dipl.-Ing. W. Hsu



Date of Issue  
(day/mo/yr)  
20/11/2017

# Certificate



Certificate no.

TA 50393494 02

**License Holder:**

MEAN WELL Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248  
Taiwan

**Manufacturing Plant:**

MEAN WELL Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248  
Taiwan

**Test report no.:** ZTW1- 50095166 001

**Client Reference:** FL-01082017-NMS240

**Tested to:** IEC 60601-1:2005+A1  
EN 60601-1:2006+A11+A1+A12

**Certified Product:** (Medical Type Switching Power Supply)

**License Fee - Units**

as page 01

Addition

Type Designation	: NMS-240-12 (MW),	1
	NMS-240-24 (MW),	1
	NMS-240-48 (MW)	1
Outputs	: DC 12V/20A (for NMS-240-12),	
	DC 24V/10A (for NMS-240-24),	
	DC 48V/5A (for NMS-240-48)	

This certificate is valid for maximum 5 years.

Appendix: 1

3

**Licensed Test mark:**



**TÜV Rheinland Taiwan Ltd**

Signature

Dipl.-Ing. W. Hsu



**Date of Issue**  
(day/mo/yr)

20/11/2017

# Certificate



Certificate no.

TA 50393494 03

**License Holder:**

MEAN WELL Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248  
Taiwan

**Manufacturing Plant:**

SuZhou MEAN WELL Technology  
Co., Ltd.  
No. 77, Jian-min Road, Dong-qiao, Pan-  
yang Ind. Park, Huang-dai Town, Xiang-  
cheng District Suzhou, Jiangsu 215152  
P. R. China

Test report no.: ZTW1- 50095166 001

Client Reference: FL-01082017-NMS240

Tested to: IEC 60601-1:2005+A1  
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd  
Signature

Dipl.-Ing. W. Hsu



Date of Issue  
(day/mo/yr)  
20/11/2017

# Certificate



Certificate no.

TA 50393494 04

**License Holder:**

MEAN WELL Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248  
Taiwan

**Manufacturing Plant:**

MEAN WELL (GUANGZHOU) ELECTRONICS  
CO., LTD HUADU BRANCH  
No. 11 Jingu South Road,  
Huadong Town, Huadu District,  
Guangzhou 510890  
P. R. China

**Test report no.:** ZTW1- 50095166 001

**Client Reference:** FL-01082017-NMS240

**Tested to:** IEC 60601-1:2005+A1  
EN 60601-1:2006+A11+A1+A12

**Certified Product:** (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



IEC 60601-1  
EN 60601-1

www.tuv.com  
ID 1419052881

TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue  
(day/mo/yr)

20/11/2017