

Certificate



Certificate no.

TA 50293226 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: RPX-60Y (MW)	10
X stands for	: S, D, T	1
Y stands for	: -3.3, -5, -12, -15, -24, -24CA, -48, A, B, C, D, 03	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.8-1.0A for X = S 1.4-0.9A for X = D, T	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Ambient Temperature	: 40°C	
Protection Class	: I	

Continued on page 02

12

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue
(day/mo/yr)

16/09/2014

Certificate



Certificate no.

TA 50293226 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Continuation

Remarks:

1. The risk management requirements of the standard were not addressed.
2. Model RPS-60-24CA is evaluated for operating at altitude 6000 m. Other models are evaluated for operating at altitude 3000 m.
3. This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)
16/09/2014

Certificate



Certificate no.

TA 50293226 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Type Designation	: RPS-40-18.2 (MW)	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.4-0.8A	
Output Voltage	: DC +18.2V	
Output Current	: 2.2A	

This certificate is valid for maximum 5 years.

1

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)
16/09/2014

Certificate



TÜV Rheinland

This page was cancelled due to factory changed.
Please check the last page for description.

Certificate no.

TA 50293226 04

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well (GUANGZHOU) Electronics
Co Ltd
2nd Floor, No. A Building,
Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou
P.R. China

Test report no.: ZTW1-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

16/09/2014

Certificate



Certificate no.

TA 50293226 05

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou Mean Well Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

16/09/2014

Certificate



Certificate no.

TA 50293226 06

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Youngyear Electronics Co., Ltd.
Jin Zhu Industries Area
Qing Xi Town, Dongguan City
Guandong
P.R. China

Test report no.: ZTWI-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

16/09/2014

Medical

台灣揚益電子股份有限公司
Taiwan Youngyear Electronics Co., Ltd.

簡仁智 先生/Mr. Tony Jean

330 桃園市林森路 113 號
113 Lin Sen Road, Taoyuan 330 Taiwan

明緯企業股份有限公司 MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd., New Taipei City 24891 Taiwan
台灣 24891 新北市五股區五權三路 28 號

Date:
9 April, 2020
Our ref.: WH/THY
Client No.: 2122906/
1220269

REGISTERED LETTER

Confirmation of License Cancellation

Dear Mr. Jean

Referring to your cancellation request on 31 March, 2020 via e-mail, we herewith confirm the cancellation of the following license sheets:

Certificate no. _ page	Service
TA 50293226 _page 6	Cancel factory Youngyear Electronics Co., Ltd. 1228582

If the cancellation of a license is submitted by the 15th of November of a year, no annual fee will be charged for the following year.

Please contact our office for safety approval of your new products.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

Yours sincerely,

TÜV RHEINLAND TAIWAN LTD.
Medical



Dipl.-Ing. W. Hsu
Certifier



Joanna Ting
Senior Executive

TÜV Rheinland Taiwan Ltd.

11F., No.758, Sec. 4, Bade Rd., Songshan Dist., Taipei City 105, Taiwan, R.O.C
Tel. (02) 2172 7000
Fax (02) 2172 1322

NEW TAIPEI BRANCH:
No. 458-18 & 458-19, Sec. 2, Fenliao Rd., Linkou Dist., New Taipei City 244, Taiwan, R.O.C
Tel. (02) 2172 1000
Fax (02) 2172 1322

TAICHUNG BRANCH:
No. 9, Ln 36, Minsheng Rd., Sec. 3, Daya Dist., Taichung City 428, Taiwan, R.O.C
Tel. (04) 2521 8888
Fax (04) 2521 8899

KAOHSIUNG BRANCH:
27F-3, No. 80, Minzu 1st Rd., Sanmin Dist., Kaohsiung City 807, Taiwan, R.O.C.
Tel. (07) 262 6000
Fax (07) 380 0577

Hotline: (02) 2172 1001
E-Mail: service-gc@tuv.com
<http://www.tuv.com>

Certificate



Certificate no.

TA 50293226 07

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10048037 002

Client Reference: FL-20150706-A12

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

27/10/2015



Dipl.-Ing. (FH) A. Klinker

明緯企業股份有限公司
MEAN WELL Enterprises Co., Ltd.

Mr. Johnny Huang, Safety & EMC Engineer
R & D Dept.

248 新北市五股區五權三路 28 號
No. 28, Wuquan 3rd Rd.
Wugu Dist., New Taipei City 248
Taiwan

Date: 7 November,
2016
Client No.: 331921
Our ref.: RNK/EVC

Ref.: Change of Factory Data

Dear Mr. Huang,

Please find the acknowledgement of the change of factory data for the existing product certificates from TÜV Rheinland Taiwan Ltd.

Delete factory:

MEAN WELL (GUANGZHOU) Electronics Co., Ltd.
2nd Floor, No. A Building, Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou, P. R. China

Additional factory:

**MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD
HUADU BRANCH**
No. 11 Jingu South Road, Huadong Town,
Huadu District, Guangzhou 510890, P.R. China

Please notes that the updated factory data will be indicated in the product certificate to be issued, started from date of above; but the existing product certificates are not to be revised with this regards.

Sincerely yours,

TÜV RHEINLAND TAIWAN LTD.
Power Supply & Display Product Line


Robert Kong
Department Manager

TÜV Rheinland Taiwan Ltd.

11F., No.758, Sec. 4, Bade
Rd., Songshan Dist., Taipei
City 105, Taiwan R.O.C.
Tel. (02) 2172 7000
Fax (02) 2528 0018
<http://www.twn.tuv.com>

TAICHUNG BRANCH:
No. 9, Ln 36, Minsheng Rd.,
Sec. 3, Daya District,
Taichung City 428,
Taiwan, R.O.C.
Tel. (04) 2560 2998
Fax (04) 2566-3598

Customer Hotline:
(02)2528 7007
www.tuv.com

MEAN WELL Enterprises Co., Ltd.
Mr. Frank Lin, Safety Engineer

Date : 08.02.2021
Our ref. : ZTW1
Your ref.: P00158594

No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Ref : TA License

Type of Equipment : Switching Power Supply
Model Designation : See Certificate
Certificate No. : TA 50293226 0008
Report No. : 10048037 004

Dear Mr. Frank Lin,

The above specified technical equipment has been tested and found to be in accordance with the relevant requirements.

Enclosed you will find the certificate of approval.

If cancellation of the certificate is submitted by November 15 in a given year, no fee will be charged for the following year.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

Please contact our office for approval of your new products.

With kind regards,

Certification Body



Dipl.-Ing. W. Hsu

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan, R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322
<http://www.tuv.com>

NEW TAIPEI BRANCH:
No. 458-18 & 458-19,
Sec. 2, Fenliao Rd., Linkou Dist.,
New Taipei City 244,
Taiwan, R. O. C.
Tel. (02) 2172-1000
Fax (02) 2172-1322

TAICHUNG BRANCH:
3F., No. 497, Zhongming South Rd.,
West Dist., Taichung City 403,
Taiwan, R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322

KAOHSIUNG BRANCH:
27F-3, No.80, Minzu 1st Rd.,
Sanmin Dist., Kaohsiung City 807,
Taiwan, R. O. C.
Tel. (07) 262-6000
Fax (07) 380-0577

Certificate



Certificate no.

TA 50293226 08

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Test report no.: ZTW1-10048037 004

Client Reference: P00158594

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Test Requirement : see above

Change

License Holder : see above

Factory : see above

Remark: 1. The risk management requirements of the standard
were not addressed.
2. This certificate is valid for 5 maximum years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.
Signature

W. Hsu
Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)
08/02/2021

Certificate



Certificate no.

TA 50293226 09

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD. HUADU BRANCH
No. 11, Jingu South Road, Huadong Town,
Huadu District, Guangzhou 510890
P.R. China

Test report no.: ZTWI- 10048037 004

Client Reference: P00158594

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

Remark: 1. The risk management requirements of the standard
were not addressed.

2. This certificate is valid for 5 maximum years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

08/02/2021

Certificate



Certificate no.

TA 50293226 10

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Test report no.: ZTW1-RC 10048037 005

Client Reference: P00984755

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

Remark: 1.The risk management requirements of the standard were not addressed.
2.Model RPS-60-24CA is evaluated for operating at altitude 6000 m. Other models are evaluated for operating at altitude 3000m.
3.This certificate is valid for maximum 5 years from last page.

Appendix: 2

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

27/11/2023

Dipl.-Ing. (FH) A. Klinker



Certificate



Certificate no.

TA 50293226 11

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

Suzhou MEAN WELL Technology Co., Ltd.
No.269, Changping Road, Huangdai Town,
Xiangcheng District, Suzhou
Jiangsu
P.R. China

Test report no.: ZTW1-RC 10048037 005

Client Reference: P00984755

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

Remark: This certificate is valid for maximum 5 years from last page.

Appendix: 2

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.


Signature

Date of Issue

(day/mo/yr)

27/11/2023

Dipl.-Ing. (FH) A. Klinker



Certificate



Certificate no.

TA 50293226 12

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD.
No. 11,
Jingu South Road, Huadu District,
510890 Guangzhou,
P.R. China

Test report no.: ZTW1-RC 10048037 005

Client Reference: P00984755

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

Remark: This certificate is valid for maximum 5 years from
last page.

Appendix: 2

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. (FH) A. Klinker



Date of Issue

(day/mo/yr)

27/11/2023